Amplite[™] Fluorimetric NADH Assay Kit *Red Fluorescence*

| Ordering Information | Storage Conditions | Instrument Platform |
|------------------------------------|--|---------------------------------|
| Product Number: 15261 (400 assays) | Keep in freezer Avoid exposure to light | Fluorescence microplate readers |

Introduction

Nicotinamide adenine dinucleotide (NAD+) and nicotinamide adenine dinucleotide phosphate (NADP+) are two important cofactors found in cells. NADH is the reduced form of NAD+, the oxidized form of NADH. NAD forms NADP with the addition of a phosphate group to the 2' position of the adenyl nucleotide through an ester linkage.

The traditional NAD/NADH and NADP/NADPH assays are done by monitoring the changes in NADH or NADPH absorption at 340 nm. The short UV wavelength of the traditional NAD/NADH and NADP/NADPH assays makes these methods to suffer low sensitivity and high interference. Due to the weak absorption of NAD and NADH, the UV absorption method requires large sample sizes, making the same NAD and NADH measurement unpractical if the availability of samples is limited.

NADH + Amplite™ NADH SensorNAD + Red Fluorescent(Non-fluorescent)Product

This AmpliteTM Fluorimetric NADH Assay Kit provides a convenient method for the detection of NADH. The enzymes in the system specifically recognize NADH in an enzyme recycling reaction. In addition, this assay has very low background since it is run in the red visible range that significantly reduces the interference resulted from biological samples. The AmpliteTM Fluorimetric NADH Assay Kit provides a sensitive, one-step assay to detect as little as 100 pico-moles of NADH in a 100 µL assay volume (1 µM). The assay can be performed in a convenient 96-well or 384-well microtiter-plate format. Its signal can be easily read by either a fluorescence microplate reader at Ex/Em = 540/590 nm or an absorbance microplate reader at ~576 nm.

Kit Key Features

| Broad Application: Sensitive: | Can be used for quantifying NADH in solutions and in cell extracts. Detect as low as 1 µM of NADH in solution. |
|----------------------------------|---|
| Continuous: | Easily adapted to automation without a separation step. |
| Convenient: | Formulated to have minimal hands-on time. No wash is required. |
| Non-Radioactive: | No special requirements for waste treatment. |

Kit Components

| Components | Amount |
|--|--------------------------------|
| Component A: NADH Recycling Enzyme Mixture | 2 bottles (lyophilized powder) |
| Component B: NADH Assay Buffer | 1 bottle (20 mL) |
| Component C: NADH Standard | 1 vial (142 µg) |

Assay Protocol for One 96-Well Plate

Brief Summary

Prepare NADH reaction mixture (50 µL) → Add NADH standards or test samples (50 µL) → Incubate at room temperature for 15 minutes – 2 hours → Monitor fluorescence intensity at Ex/Em = 540/590 nm

Note: Thaw one of each kit component at room temperature before starting the experiment.

1. Prepare NADH stock solution:

Add 200 μL of PBS buffer into the vial of NADH standard (Component C) to make 1 mM (1 nmol/ μL) NADH stock solution.

Note: The unused NADH stock solution should be divided into single use aliquots and stored at -20°C.

2. Prepare NADH reaction mixture:

Add 10 mL of Amplite[™] NADH Assay Buffer (Component B) to the bottle of NADH Recycling Enzyme Mixture (Component A), and mix well.

Note: This NADH reaction mixture is enough for two 96-well or four 384-well plates. The unused NADH reaction mixture should be divided into single use aliquots and stored at -20°C.

3. Prepare serial dilutions of NADH standard (0 to 100 μ M):

- 3.1 Add 50 μL of NADH stock solution (from Step 1) into 450 μL PBS buffer (pH 7.4) to generate 100 μM (100 pmol/μL) NADH standard solution. Note: Diluted NADH standard solution is unstable, and should be used within 4 hours.
- 3.2 Take 200 μ L of 100 μ M NADH standard solution to perform 1:3 serial dilutions to get 30, 10, 3, 1, 0.3, 0.1 and 0 μ M serial dilutions of NADH standard.
- 3.3 Add serial dilutions of NADH standard and NADH containing test samples into a solid black 96-well microplate as described in Tables 1 and 2 Note: Prepare cells or tissue samples as desired.

| | | | | 1 | 1 | | | |
|------------|-----|----|----|---|---|--|--|--|
| BL | BL | TS | TS | | | | | |
| NS1 | NS1 | | | | | | | |
| NS2 | NS2 | | | | | | | |
| NS3 | NS3 | | | | | | | |
| NS4 | NS4 | | | | | | | |
| NS4 NS5 | NS5 | | | | | | | |
| NS6 | NS6 | | | | | | | |
| NS7 | NS7 | | | | | | | |

Table 1 Layout of NADH standards and test samples in a solid black 96-well microplate

Note: NS= *NADH Standards, BL*=*Blank Control, TS*=*Test Samples.*

 Table 2 Reagent composition for each well

| NADH Standard | Blank Control | Test Sample |
|--------------------------|---------------|-------------|
| Serial Dilutions*: 50 µL | PBS: 50 μL | 50 μL |

*Note: Add the serially diluted NADH standards from 0.1 μ M to 100 μ M into wells from NS1 to NS7 in duplicate.

4. Run NADH assay in supernatants reaction:

- 4.1 Add 50 μL of NADH reaction mixture (from Step 2) into each well of NADH standard, blank control, and test samples (see Step 3.3) to make the total NADH assay volume of 100 μL/well Note: For a 384-well plate, add 25 μL of sample and 25 μL of NADH reaction mixture into each well.
- 4.2 Incubate the reaction at room temperature for 15 minutes to 2 hours, protected from light.

4.3 Monitor the fluorescence increase with a fluorescence plate reader at Ex/Em = 530 - 570/590 - 600 nm (optimal at Ex/Em = 540/590 nm).

Note: The contents of the plate can also be transferred to a white clear bottom plate and read by an absorbance microplate reader at the wavelength of 576 \pm 5 nm. The absorption detection has lower sensitivity compared to fluorescence reading.

Data Analysis

The fluorescence in blank wells (with the PBS buffer only) is used as a control, and is subtracted from the values for those wells with the NADH reactions. A NADH standard curve is shown in Figure 1. *Note: The fluorescence background increases with time, thus it is important to subtract the fluorescence intensity value of the blank wells for each data point.*




Figure 1. NADH dose response was measured with AmpliteTM NADH Assay Kit in a 96-well black plate using a NOVOStar microplate reader (BMG Labtech). As low as 1 μ M (100 pmols/well) of NADH can be detected with 1 hour incubation (n=3) while there is no response from NAD.

<u>References</u>

- 1. Ziegenhorn J, Senn M, Bucher T. (1976) Molar absorptivities of beta-NADH and beta-NADPH. Clin Chem, 22, 151.
- 2. Ikegami T, Kameyama E, Yamamoto SY, Minami Y, Yubisui T. (2007) Structure and Properties of the Recombinant NADH-Cytochrome b(5) Reductase of Physarum polycephalum. Biosci Biotechnol Biochem.
- 3. Kimura N, Fukuwatari T, Sasaki R, Shibata K. (2006) Comparison of metabolic fates of nicotinamide, NAD+ and NADH administered orally and intraperitoneally; characterization of oral NADH. J Nutr Sci Vitaminol (Tokyo), 52, 142.
- 4. O'Donnell JM, Kudej RK, LaNoue KF, Vatner SF, Lewandowski ED. (2004) Limited transfer of cytosolic NADH into mitochondria at high cardiac workload. Am J Physiol Heart Circ Physiol, 286, H2237.

Warning: This kit is only sold to end users. Neither resale nor transfer to a third party is allowed without written permission from AAT Bioquest. Chemical analysis of the kit components is strictly prohibited. Please call us at 408-733-1055 or e-mail us at info@aatbio.com if you have any questions.