



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

Corporate Headquarters
AMRESCO, LLC
6681 Cochran Road
Solon, Ohio USA 44139-4300
440/349-1199 FAX: 440/349-1182
www.amresco-inc.com
Email: info@amresco-inc.com

CERTIFICATE OF QUALITY

Product: NEXT GEL, 15%
Grade: PROTEOMICS GRADE
Code: M258
Appearance: Clear, colorless liquid.
Mol. Formula: Not Applicable
Mol. Weight: Not Applicable

Description:

<u>Assay:</u>	<u>Min</u>	<u>Max</u>	<u>Units</u>	<u>Disposition</u>
Conductivity (1:20, Water)	430	550	uS	PASS
Electrophoresis	PASS		P/F	PASS
pH @25C	6.8	7.2		PASS
Polymerization Time		30	minutes	PASS
Protease		NONE	P/F	PASS

Storage: Store at room temperature. Protect from moisture.

m258ptm

This signature indicates that the above material has met all quality specifications and has been reviewed by a quality representative.

Signature:

Title:

QC MANAGER

Date:

6/13/2012