



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

Corporate Headquarters
AMRESCO, LLC
6681 Cochran Road
Solon, Ohio USA 44139-4300
440/349-1199 FAX: 440/349-1182
www.amresco-inc.com
Email: info@amresco-inc.com

CERTIFICATE OF QUALITY

Product: RIPA LYSIS BUFFER
Grade: BIOTECHNOLOGY GRADE
Code: N653
Appearance: Clear, colorless solution.
Mol. Formula: Not Applicable
Mol. Weight: Not Applicable

Description:

<u>Assay:</u>	<u>Min</u>	<u>Max</u>	<u>Units</u>	<u>Disposition</u>
pH @25C	7.4	7.6		PASS
Tris Titration	48	55	mM	PASS

Storage: Store cold. Keep dry. Warm to room temperature before opening.

n653bio

This signature indicates that the above material has met all quality specifications and has been reviewed by a quality representative.

Signature:

Title:

QC MANAGER

Date:

6/22/2012