

ELISA for Monkey Apolipoprotein E

Product Code: 3712M-1A-20

CONTENTS, development kit for 20 plates:

Vial 1 (red top)

Monoclonal antibody E981 (1 ml)

Concentration: 0.5 mg/ml

Vial 2 (yellow top)

Biotinylated monoclonal antibody E887 (500 µl)

Concentration: 0.5 mg/ml

Vial 3 (white top)

Streptavidin-Alkaline Phosphatase (250 µl)

Vial 4

Recombinant apoE3 standard (5 µg)

To ensure total recovery of stated quantity, vials have been overfilled.

STORAGE:

Shipped at ambient temperature. On arrival box 1 should be stored refrigerated at 4-8°C and box 2 should be stored frozen at -20°C.

General


Intended use: For quantitative determination of native and recombinant Apolipoprotein E in serum/plasma samples and cell culture supernatants.

Serum/plasma samples: The mAbs will recognize apoE only in the presence of non-ionic detergents at a concentration of 0.01-0.5%. Avoid vortex in the presence of detergent. We recommend the use of Assay buffer (product code: 3652-J2) for dilution of samples, standard and detection antibody. The buffer also prevents false positive read-outs which may be caused by interference of heterophilic antibodies commonly found in plasma and serum. Serum/plasma samples containing EDTA, citrate or heparin may be used. However, heparin containing samples will give higher apoE values due to displacement of proteoglycan bound apoE. Please contact Mabtech for further information.

Reagents: Antibodies are supplied in sterile-filtered (0.2 μ m) PBS with sodium azide (0.02%). Streptavidin-ALP is supplied in 0.1 M Tris buffer with 0.15% Kathon CG.

Standard range: 0.1-10 ng/ml

Standard calibration: No international standard exists for calibration.



Guidelines for Monkey Apolipoprotein E ELISA

- Day 1**
1. Coat a high protein binding ELISA plate with mAb E981, diluted to 2 µg/ml in PBS, pH 7.4, by adding 100 µl/well. Incubate overnight at 4-8°C.
- Day 2**
2. Wash twice with PBS (200 µl/well).
 3. Block plate by adding 200 µl/well of PBS with 0.05% Tween 20 containing 0.1% BSA (incubation buffer). Incubate for 1 hour at room temperature.
 4. Wash five times with PBS containing 0.05% Tween.
 5. Prepare apoE standard by reconstituting contents of vial 4 in 1 ml PBS with 0.5 mM DTT and 0.1% BSA, do not stir. It is important to wait 20 minutes before resuspending the liquid. This gives a stock solution of 5 µg/ml which should be used immediately or stored in aliquots at -20°C for future use. The recommended standard dilutions range from 0.1-10 ng/ml.
 6. Add 100 µl/well of samples or standards diluted in incubation buffer or Assay buffer for serum/plasma samples and incubate for 1 to 2 hours at room temperature.
 7. Wash as in step 4.
 8. Add 100 µl/well of mAb E887-biotin at 1 µg/ml in incubation buffer or Assay buffer for serum/plasma samples. Incubate for 1 hour at room temperature.
 9. Wash as in step 4.
 10. Add 100 µl/well of Streptavidin-ALP diluted 1:1000 in incubation buffer. Incubate for 1 hour at room temperature.
 11. Wash as in step 4.
 12. Add 100 µl/well of appropriate substrate solution e.g. p-nitrophenyl-phosphate (pNPP).
 13. Measure the optical density (405 nm for pNPP) in an ELISA reader after suitable developing time.

NOTE; for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages therefrom.



MABTECH AB
Box 1233
SE-131 28 Nacka Strand
Sweden
Tel: +46 8 716 27 00
Fax: +46 8 716 27 01
E-mail: mabtech@mabtech.com
www.mabtech.com

MABTECH Inc
M.E.B. 220
3814 West Street
Cincinnati, OH 45227
USA
Tel: +1 513 871 4500
Fax: +1 513 871 7353
E-mail: mabtech.usa@mabtech.com

MABTECH AB Büro Deutschland
Germany
Tel: +49 40 4135 7935
Fax: +49 40 4135 7945
E-mail: mabtech.de@mabtech.com

2013-01-24

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the following standards:



MABTECH AUSTRALIA Pty Ltd
resolvingIMAGES
Unit 22, 196 Settlement Road
Thomastown Victoria 3074
Australia
Tel: +61 3 9466 4007
Fax: +61 3 9466 4003
E-mail: mabtech.au@mabtech.com

MABTECH AB Bureau de liaison France
BP 255, 1300 route des Crêtes
06905 Sophia Antipolis
France
Tel: +33 (0)4 92 38 80 70
Fax: +33 (0)4 92 38 80 71
E-mail: mabtech.fr@mabtech.com