

# ELISA for Mouse IgG

Product Code: 3825-1AD-6

---

## **CONTENTS:**

### **Vial 1 (yellow top)**

Anti-IgG antibody (150  $\mu$ l)

Concentration: 0.5 mg/ml

### **Vial 2 (green top)**

ALP-conjugated anti-IgG antibody (80  $\mu$ l)

### **Vial 3**

Lyophilised mouse IgG standard

To ensure total recovery of stated quantity, vials have been overfilled.

## **STORAGE:**


Shipped at ambient temperature. On arrival box 1 should be stored refrigerated at 4-8°C and box 2 should be stored frozen at -20°C.

# General

**Intended use:** For quantitative determination of native mouse IgG in serum and plasma.

**Reagents:** Anti-IgG antibody is supplied in sterile-filtered (0.2  $\mu\text{m}$ ) PBS with sodium azide (0.02%). ALP-conjugated anti-IgG antibody is supplied in 0.1 M Tris-buffer with 0.15% Kathon CG.

**Recommended standard dilution:** 0.1-500 ng/ml



# Guidelines for Mouse IgG ELISA

- Day 1**
1. Coat a high protein binding ELISA plate with anti-IgG antibody, diluted to 1  $\mu\text{g}/\text{ml}$  in PBS, pH 7.4, by adding 100  $\mu\text{l}$ /well. Incubate overnight at 4-8°C.
- Day 2**
2. Wash twice with PBS (200  $\mu\text{l}$ /well).
  3. Block plate by adding 200  $\mu\text{l}$ /well of PBS with 0.05% Tween 20 (PBS-Tween) containing 0.1% BSA (incubation buffer\*). Incubate for 1 hour at room temperature.
  4. Wash five times with PBS-Tween.
  5. Prepare mouse IgG standard by reconstituting contents of vial 3 in 500  $\mu\text{l}$  PBS to make up a stock solution of 50  $\mu\text{g}/\text{ml}$ . The stock solution should be used immediately or stored in aliquots at -20°C for future use. We recommend the aliquots not be refrozen after initial use. For the test, prepare dilutions of the stock using the standard range as a guideline.
  6. Add 100  $\mu\text{l}$ /well of samples or standards diluted in incubation buffer and incubate for 2 hours at room temperature.
  7. Wash as in step 4.
  8. Add 100  $\mu\text{l}$ /well of anti-IgG-ALP diluted 1:1000 in incubation buffer. Incubate for 1 hour at room temperature.
  9. Wash as in step 4.
  10. Add 100  $\mu\text{l}$ /well of appropriate substrate solution e.g. p-nitrophenyl-phosphate (pNPP).
  11. Measure the optical density (405 nm for pNPP) in an ELISA reader after suitable developing time.

\* The same buffer is used for blocking and for dilution.

**NOTE; for research use only.**

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages therefrom.



MABTECH AB  
Box 1233  
SE-131 28 Nacka Strand  
Sweden  
Tel: +46 8 716 27 00  
Fax: +46 8 716 27 01  
E-mail: [mabtech@mabtech.com](mailto:mabtech@mabtech.com)  
[www.mabtech.com](http://www.mabtech.com)

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the following standards:



MABTECH Inc  
M.E.B. 220  
3814 West Street  
Cincinnati, OH 45227  
USA  
Tel: +1 513 871 4500  
Fax: +1 513 871 7353  
E-mail: [mabtech.usa@mabtech.com](mailto:mabtech.usa@mabtech.com)

MABTECH AUSTRALIA Pty Ltd  
44 Gresswell Road  
Macleod, VIC 3085  
Australia  
Tel: +61 3 9459 9630  
Fax: +61 3 9455 0084  
E-mail: [mabtech.au@mabtech.com](mailto:mabtech.au@mabtech.com)

MABTECH AB Büro Deutschland  
Germany  
Tel: +49 40 4135 7935  
Fax: +49 40 4135 7945  
E-mail: [mabtech.de@mabtech.com](mailto:mabtech.de@mabtech.com)

MABTECH AB Bureau de liaison France  
BP 255, 1300 route des Crêtes  
06905 Sophia Antipolis  
France  
Tel: +33 (0)4 92 38 80 70  
Fax: +33 (0)4 92 38 80 71  
E-mail: [mabtech.fr@mabtech.com](mailto:mabtech.fr@mabtech.com)

2012-10-25