

## Technical Data Sheet

## Mouse IL-10 Standard

## Product Information

Material Number:	558474
Component:	51-9003528
Description:	Mouse IL-10 Standard
Size:	1 Vial(s) (1 ea)
Storage Buffer:	Lyophilized in an aqueous buffered solution containing BSA and ProClin™ 150.

## Description

This product contains lyophilized recombinant mouse Interleukin-10 (IL-10). When reconstituted, this product can be used as a standard control with the BD™ CBA Mouse IL-10 Flex Set (Cat. No. 558300) or with the BD™ CBA Mouse IL-10 Enhanced Sensitivity Flex Set (Cat. No. 562263). Please refer to the technical data sheet included with the BD™ CBA Mouse IL-10 Flex Set or the BD™ CBA Mouse IL-10 Enhanced Sensitivity Flex Set for further information.

## Preparation and Storage

Store the lyophilized standard at 4°C prior to reconstitution.

Lyosphere(s) from 1 vial of Mouse IL-10 Standard, when reconstituted in 4.0 mL Assay Diluent, approximates the CBA activity of Mouse IL-10 at a concentration of 2500 pg/mL. Following reconstitution, maintain the Mouse IL-10 at 4°C and use within 12 hours. Any unused reconstituted standard must be discarded after this period (do not store or reuse).

## Application Notes

## Application

Flow cytometry	Routinely Tested
----------------	------------------

## Recommended Assay Procedure:

The Mouse IL-10 Standard has been tested to assure to function as a standard when used with the BD™ Cytometric Bead Array (CBA). Investigators are encouraged to refer to the technical data sheet for the BD™ CBA Mouse IL-10 Flex Set (Cat. No. 558300) or the BD™ CBA Mouse IL-10 Enhanced Sensitivity Flex Set (Cat. No. 562263). The Mouse IL-10 Standard is lyophilized and should be reconstituted and serially diluted before mixing with CBA capture beads and PE detection reagents. An abbreviated instruction has been provided below (Preparation of Mouse IL-10 Standard Protocol). The Mouse IL-10 Standard should be reconstituted using Assay Diluent, which is provided with the BD™ CBA Mouse/Rat Soluble Protein Master Buffer Kit (Cat. No. 558266 or 558267), the BD™ CBA Mouse Enhanced Sensitivity Master Buffer Kit (Cat. No. 562246 or 562248) or it may also be purchased separately (Cat. No. 560104).

## Preparation of Mouse IL-10 Standard for use with the BD™ CBA Mouse IL-10 Flex Set (Cat. No. 558300):

1. Open one vial of the lyophilized Mouse IL-10 Standard. Transfer the lyosphere(s) to a polypropylene tube (BD Falcon™, Cat. No. 352097). Label the tube "Top Standard".
2. Reconstitute the standard with 4.0 mL of Assay Diluent. Allow the reconstituted standard to equilibrate for at least 15 minutes before making dilutions. **Mix the reconstituted protein by pipette only. Do not vortex or mix vigorously.**
3. Label additional 12 x 75 mm tubes (BD Falcon™, Cat. No. 352008) and arrange them in the following order: Top Standard, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, and 1:256.
4. Add 500 µl of Assay Diluent to each of the dilution tubes.
5. Perform a serial dilution by transferring 500 µl from the Top Standard to the 1:2 dilution tube and mix thoroughly. **Do not vortex, mix by pipet only.** Continue making serial dilutions by transferring 500 µl from the 1:2 tube to the 1:4 tube and so on to the 1:256 tube and mix thoroughly. Prepare one tube containing only Assay Diluent to serve as the 0 pg/mL negative control.

## Preparation of Mouse IL-10 Standard for use with the BD™ CBA Mouse IL-10 Enhanced Sensitivity Flex Set (Cat. No. 562263):

1. Open one vial of the lyophilized Mouse IL-10 Standard. Transfer the lyosphere(s) to a polypropylene tube (BD Falcon™, Cat. No. 352097).
2. Reconstitute the standard with 4.0 mL of Assay Diluent. Allow the reconstituted standard to equilibrate for at least 15 minutes before making dilutions. **Mix the reconstituted protein by pipette only. Do not vortex or mix vigorously.**
3. Label additional 12 x 75 mm tubes (BD Falcon™, Cat. No. 352008) and arrange them in the following order: Top Standard, 1:3, 1:9, 1:27, 1:81, 1:243, and 1:729.
4. Add 460 µl of Assay Diluent into the Top Standard tube.
5. Add 400 µl of Assay Diluent to each of the remaining dilution tubes.
6. Transfer 40 µl of the reconstituted standard into the Top Standard tube. Mix by pipette or gently vortexing for no longer than 3 seconds.

## BD Biosciences

bdbiosciences.com

United States	Canada	Europe	Japan	Asia Pacific	Latin America/Caribbean
877.232.8995	866.979.9408	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit [bdbiosciences.com/contact](http://bdbiosciences.com/contact)

*Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.*

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2015 BD

558474 Rev. 6



7. Perform a serial dilution by transferring 200 µl from the Top Standard to the 1:3 dilution tube and mix thoroughly. **Do not vortex, mix by pipet only.** Continue making serial dilutions by transferring 200 µl from the 1:3 tube to the 1:9 tube and so on to the 1:729 tube and mix thoroughly. Prepare one tube containing only Assay Diluent to serve as the 0 fg/mL negative control.

### Suggested Companion Products

Catalog Number	Name	Size	Clone
558300	Mouse IL-10 Flex Set	100 Tests	(none)
558266	Mouse/Rat Soluble Protein Master Buffer Kit	100 Tests	(none)
558267	Mouse/Rat Soluble Protein Master Buffer Kit	500 Tests	(none)
560104	Assay Diluent	30 mL	(none)
352097	BD Falcon™ Conical Centrifuge Tubes, 15 ML		(none)
352008	BD Falcon™ Round Bottom Tubes, 6 ML		(none)
562263	Mouse IL-10 Enhanced Sensitivity Flex Set	100 Tests	(none)
562246	Mouse Enhanced Sensitivity Master Buffer Kit	100 Tests	(none)
562248	Mouse Enhanced Sensitivity Master Buffer Kit	500 Tests	(none)

### Product Notices

1. ProClin is a trademark of Rohm and Haas Company.
2. Source of all serum proteins is from USDA inspected abattoirs located in the United States.
3. Please refer to [www.bdbiosciences.com/pharming/en/protocols](http://www.bdbiosciences.com/pharming/en/protocols) for technical protocols.
4. Warning: CBA lyophilized standard contains 0.02% (w/w) of a CMIT/MIT mixture (3:1), which is a mixture of: 5-chloro-2-methyl-4-isothiazolin-3-one [EC No 247-500-7] and 2-methyl-4-isothiazolin-3-one [EC No 220-239-6] (3:1).  
Hazard statement: May cause an allergic skin reaction.  
Precautionary statements: Wear protective gloves/eye protection. Wear protective clothing. Avoid breathing mist/vapours/spray. If skin irritation or rash occurs: Get medical advice/attention. IF ON SKIN: Wash with plenty of water. Dispose of contents/container in accordance with local/regional/national/international regulations.

### BD Biosciences

[bdbiosciences.com](http://bdbiosciences.com)

<b>United States</b>	<b>Canada</b>	<b>Europe</b>	<b>Japan</b>	<b>Asia Pacific</b>	<b>Latin America/Caribbean</b>
877.232.8995	866.979.9408	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit [bdbiosciences.com/contact](http://bdbiosciences.com/contact)

*Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.*

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2015 BD

558474 Rev. 6

