

Technical Data Sheet

Phospho p38 (T180/Y182) Standard

Product Information

Material Number:	560028
Component:	51-9004903
Description:	Phospho p38 (T180/Y182) Standard
Size:	1 vial (1 ea)
Storage Buffer:	Lyophilized in an aqueous buffered solution containing BSA and ProClin™ 150.

Description

This product contains lyophilized phosphorylated recombinant p38 (T180/Y182). When reconstituted, this product can be used as a standard control with the BD™ CBA Phospho p38 (T180/Y182) Flex Set (Cat. No. 560010). Please refer to the technical data sheet included with the BD™ CBA Phospho p38 (T180/Y182) Flex Set for further information.

Preparation and Storage

Store the lyophilized standard at 4°C prior to reconstitution.

Lyosphere(s) from 1 vial of Phospho p38 (T180/Y182) Standard, when reconstituted in 0.1 mL Assay Diluent, approximates the CBA activity of Phospho p38 (T180/Y182) at a concentration of 50,000 Units/mL. Following reconstitution, the Phospho p38 (T180/Y182) Standard may be stored at 4°C for up to 3 months. Any unused, reconstituted standard must be discarded after this period. When using reconstituted standard, warm to 37°C and vortex to mix thoroughly before use.

Application Notes

Application

Flow cytometry	Routinely Tested
----------------	------------------

Recommended Assay Procedure:

The Phospho p38 (T180/Y182) Standard has been tested to assure to function as a standard when used with the BD™ Cytometric Bead Array (CBA), such as the BD™ CBA Phospho p38 (T180/Y182) Flex Set (Cat. No. 560010). Investigators are encouraged to refer to the technical data sheet for the BD™ CBA Phospho p38 (T180/Y182) Flex Set (Cat. No. 560010) as well as the product manual for the BD™ CBA Cell Signaling Master Buffer Kit (Cat. No. 560005 or 560006). An abbreviated instruction has been provided below (Preparation of Phospho p38 (T180/Y182) Standard Protocol). The Phospho p38 (T180/Y182) Standard should be reconstituted using Assay Diluent, which is provided with the BD™ CBA Cell Signaling Master Buffer Kit (Cat. No. 560005 or 560006) or it may also be purchased separately (Cat. No. 560104).

Preparation of Phospho p38 (T180/Y182) Standard: The Phospho p38 (T180/Y182) Standard is lyophilized and should be reconstituted and serially diluted before mixing with CBA capture beads and PE detection reagents.

1. Open one vial of the lyophilized Phospho p38 (T180/Y182) Standard. Transfer the lyosphere(s) to a 1.5 mL microfuge tube.
2. Reconstitute the standard with 0.1 mL of Assay Diluent, warming the tube to 37°C, and vortexing. This creates a 50x stock solution.
3. Label 12 x 75 mm tubes (BD Falcon™, Cat. No. 352008) and arrange them in the following order: Top Standard, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, and 1:256.
4. Add 20 µl of the 50x Standard to one tube. Label the tube "Top Standard".
5. Add Assay Diluent to Top Standard to achieve a final volume of 1 mL and a 1x concentration.
6. Add 500 µl of Assay Diluent to each of the remaining dilution tubes.
7. Perform a serial dilution by transferring 500 µl from the Top Standard to the 1:2 dilution tube and mix thoroughly. Continue making serial dilutions by transferring 500 µl from the 1:2 tube to the 1:4 tube and so on to the 1:256 tube and mix thoroughly. Prepare one tube containing only Assay Diluent to serve as the 0 Units/mL negative control.

Suggested Companion Products

Catalog Number	Name	Size	Clone
560010	Phospho p38 (T180/Y182) Flex Set	100 tests	(none)
560005	Cell Signaling Master Buffer Kit	100 tests	(none)
560006	Cell Signaling Master Buffer Kit	500 tests	(none)
560104	Assay Diluent	30 ml	(none)
352008	BD Falcon™ Round Bottom Tubes, 6 ML		(none)

BD Biosciences

bdbiosciences.com

United States	Canada	Europe	Japan	Asia Pacific	Latin America/Caribbean
877.232.8995	800.268.5430	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.

Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2014 BD



Product Notices

- 1. ProClin is a trademark of Rohm and Haas Company.
- 2. Source of all serum proteins is from USDA inspected abattoirs located in the United States.
- 3. Please refer to www.bdbiosciences.com/pharming/protocols for technical protocols.

BD Biosciences

bdbiosciences.com					
United States	Canada	Europe	Japan	Asia Pacific	Latin America/Caribbean
877.232.8995	800.268.5430	32.2.400.98.95	0120.8555.90	65.6861.0633	55.11.5185.9995

For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.
Unless otherwise noted, BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2014 BD

