Technical Data Sheet V500 Mouse Anti-Human CD3

Product Information

Material Number:
Alternate Name:
Size:
Vol. per Test:
Clone:
Isotype:
Reactivity:
Workshop:
Storage Buffer:

561417
CD3e; CD3E; T3E; TCRE ; T-cell surface antigen T3/Leu-4 epsilon 25 tests
5 μl
UCHT1
Mouse IgG1, κ
QC Testing: Human
III 471
Aqueous buffered solution containing protein stabilizer, glycerol and ≤0.09% sodium azide.

Description

The UCHT1 monoclonal antibody specifically binds to the human CD3ε-chain, a 20-kDa subunit of the CD3/T cell antigen receptor complex. CD3ε is expressed on 70-80% of normal human peripheral blood lymphocytes and 60-85% of thymocytes. Studies from the HLDA Workshop show that this antibody is mitogenic for CD3ε-positive cells when used in conjunction with costimulatory agents such as pokeweed mitogen or anti-CD28 antibody. CD3 plays a central role in signal transduction during antigen recognition. The UCHT1 antibody stains both surface and intracellular CD3ε unlike the other CD3 clone, HIT3a, that stains only extracellular CD3ε.

The antibody is conjugated to BD Horizon[™] V500, which has been developed for use in multicolor flow cytometry experiments and is available exclusively from BD Biosciences. It is excited by the Violet laser with an Ex max of 415 nm and Em Max at 500 nm. BD Horizon V500 conjugates emit at a similar wavelength to Amcyan yet exhibit reduced spillover into the FITC channel. For more information on BD Horizon V500, visit bdbiosciences.com/colors.

When compensating dyes in this spectral range (such as Horizon[™] V500 and AmCyan), the most accurate compensation can be obtained using single stained cellular controls. Due to spectral differences between cells and beads in this channel, using BD CompBeads can result in spillover errors for V500 and AmCyan reagents. Therefore, the use of BD CompBeads or BD CompBeads Plus to determine spillover values for these reagents is not recommended. Different V500 reagents (e.g. CD4 vs. CD45) can have slightly different fluorescence spillover therefore, it may also be necessary to use clone specific compensation controls when using these reagents.



Flow cytometric analysis of CD3 expressed on human peripheral lymphocytes. Human whole blood was stained with the BD Horizon™ V500 Mouse anti-Human CD3 antibody (Cat. No. 561416/561417; solid line histogram) or with a BD Horizon™ V500 Mouse IgG1, κ Isotype Control (Cat. No. 560787; dashed line histogram). The erythrocytes were lysed with BD PharmLyse™ Lysing Buffer (Cat. No. 555899). The fluorescence histograms were derived from events with the forward and side light-scatter characteristics of viable lymphocytes. Flow cytometry was performed using a BD™ LSR II Flow Cytometer System.

BD Biosciences

bdbiosciences.com United States Canada Asia Pacific Latin America/Caribbean Europe Japan 32.53.720.550 0120.8555.90 877.232.8995 888.268.5430 65.6861.0633 0800.771.7157 For country-specific contact information, visit bdbiosciences.com/how_to_order/ Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton Dickinson and Company is strictly prohibited. For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale. BD, BD Logo and all other trademarks are the property of Becton, Dickinson and Company. ©2011 BD



Preparation and Storage

The monoclonal antibody was purified from tissue culture supernatant or ascites by affinity chromatography. The antibody was conjugated with BD Horizon[™] V500 under optimum conditions, and unreacted BD Horizon[™] V500 was removed. Store undiluted at 4°C and protected from prolonged exposure to light. Do not freeze.

Application Notes

Application	
-------------	--

	Flow cytometry	Routinely Tested		
Suggested Companion Products				

Catalog Number	Name	Size	Clone
555899	Lysing Buffer	100 ml	(none)
560787	V500 Mouse IgG1, κ Isotype Control	0.1 mg	X40
554656	Stain Buffer (FBS)	500 ml	(none)
561416	V500 Mouse Anti-Human CD3	100 tests	UCHT1

Product Notices

- 1. This reagent has been pre-diluted for use at the recommended Volume per Test. We typically use 1×10^{6} cells in a 100-µl experimental sample (a test).
- 2. An isotype control should be used at the same concentration as the antibody of interest.
- 3. Please refer to www.bdbiosciences.com/pharmingen/protocols for technical protocols.
- 4. Caution: Sodium azide yields highly toxic hydrazoic acid under acidic conditions. Dilute azide compounds in running water before discarding to avoid accumulation of potentially explosive deposits in plumbing.
- 5. For fluorochrome spectra and suitable instrument settings, please refer to our Fluorochrome Web Page at www.bdbiosciences.com/colors.
- 6. BD Horizon[™] V500 has a maximum absorption of 415 nm and maximum emission of 500 nm. Before staining with this reagent, please

confirm that your flow cytometer is capable of exciting the fluorochrome and discriminating the resulting fluorescence.

References

Barclay NA, Brown MH, Birkeland ML, et al, ed. *The Leukocyte Antigen FactsBook*. San Diego, CA: Academic Press; 1997. (Biology) Beverley PC, Callard RE. Distinctive functional characteristics of human "T" lymphocytes defined by E rosetting or a monoclonal anti-T cell antibody. *Eur J Immunol.* 1981; 11(4):329-334. (Biology)

Ernst DN, Shih CC. CD3 complex. J Biol Regul Homeost Agents. 2000; 14(3):226-229. (Biology)

Knapp W, Dorken B, Rieber EP, et al, ed. Leucocyte Typing IV. New York: Oxford University Press; 1989:1-1208. (Clone-specific)

Lanier LL, Allison JP, Phillips JH. Correlation of cell surface antigen expression on human thymocytes by multi-color flow cytometric analysis: implications for differentiation. J Immunol. 1986; 137(8):2501-2507. (Biology)

McMichael AJ, Beverly PCL, Gilks W, et al, ed. Leukocyte Typing III: White Cell Differentiation Antigens. New York: Oxford University Press; 1987. (Clone-specific)

Schlossman S, Boumell L, et al, ed. Leucocyte Typing V. New York: Oxford University Press; 1995. (Biology)

Schlossman SF, Boumsell L, Gilks W, et al, ed. *Leukocyte Typing V: White Cell Differentiation Antigens*. New York: Oxford University Press; 1995. (Clone-specific)