

C Tubes (25 tubes)	130-093-237
C Tubes (100 tubes)	130-096-334

Contents

1. Description

- 1.1 Background information
- 1.2 Technical specifications
- 1.3 Reagent and instrument requirements
- 1.4 Applications

1. Description

Components	C Tubes (25 tubes), sterile and single packed or C Tubes (100 tubes), sterile packed as 4×25 tubes.
Storage	Store tubes dry and protected from light at room temperature (19–25 °C). The expiration date is indicated on the box label. Do not use after this date.

1.1 Background information

The purple-colored gentleMACS™ C Tubes are exclusively designed for automated dissociation using the gentleMACS Dissociators. C Tubes are used for very gentle and efficient dissociation of tissues to obtain single-cell suspensions. The cap of the C Tubes bears a specially designed stator and a rotor with a septum-sealed opening in the center. The sample material and medium or buffer is placed in the opened C Tubes. The tubes are tightly closed with the cap. The tube is attached upside down to the sleeve of the gentleMACS Dissociators. It has to be ensured that the sample material is located in the area of the rotor/stator. The dissociation process is started by running an appropriate gentleMACS Program. After processing, the C Tube is detached and the dissociated material is removed from the C Tube.

1.2 Technical specifications

- Sample sizes
 - 500 µL – maximal 10 mL total volume
 - 20–4000 mg of tissue

▲ **Note:** The maximal amount of tissue depends on the consistency of the tissue as well as on the respective gentleMACS Protocol.

▲ **Note:** Dependent on the consistency of the tissue it might be recommended to cut the sample into smaller pieces before processing on the gentleMACS Dissociators.

▲ **Note:** The tubes are suitable for a broad range of tissues, for example, spleen, liver, or brain. However, very hard material, such as bone, may damage the tubes.
- Dimensions
 - Length: 87 mm
 - Diameter: 30 mm

- Recommended working temperature is between 4 °C and 40 °C.
- Maximal centrifugal acceleration
Relative centrifugal force (RCF) = 3000×g
- The tubes are for single use only.
- To close the C Tubes, turn the cap clockwise beyond the first resistance.

1.3 Reagent and instrument requirements

- gentleMACS Dissociator (# 130-093-235), gentleMACS Octo Dissociator (# 130-095-937), or gentleMACS Octo Dissociator with Heaters (# 130-096-427)
- For removal of dissociated material from the closed C Tubes, special pipette tips of extended length are required, e.g., ART® 1000 REACH™ pipet tips (Molecular BioProducts, Inc.).
 - ▲ **Note:** Regular 1000 µL pipette tips are not suitable for pipetting via the septum. Volumes of up to 200 µL can be added via the septum by using regular pipette tips.
- (Optional) MACS® Tissue Dissociation Kits
- (Optional) Pre-Separation Filters or MACS SmartStrainers to remove cell clumps.
- (Optional) MACS Cell Separation Reagents
- (Optional) MACS Antibodies for cell analysis

1.4 Applications

Usually, C Tubes are designed for the dissociation of tissues to obtain single-cell suspensions.

For further details and more applications, please refer to www.miltenyibiotec.com. The gentleMACS Protocols are also online available at www.miltenyibiotec.com/protocols.

All protocols and data sheets are available at www.miltenyibiotec.com.

Warranty

The products sold hereunder are warranted only to be free from defects in workmanship and material at the time of delivery to the customer. Miltenyi Biotec GmbH makes no warranty or representation, either expressed or implied, with respect to the fitness of a product for a particular purpose. There are no warranties, expressed or implied, which extend beyond the technical specifications of the products. Miltenyi Biotec GmbH's liability is limited to either replacement of the products or refund of the purchase price. Miltenyi Biotec GmbH is not liable for any property damage, personal injury or economic loss caused by the product.

MACS is a registered trademark and gentleMACS is a trademark of Miltenyi Biotec GmbH.

ART is a registered trademark and REACH is a trademark of Molecular BioProducts, Inc.

Unless otherwise specifically indicated, Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.

Copyright © 2014 Miltenyi Biotec GmbH. All rights reserved.